

TECHNICAL DATA

Fluke BP291 4800 mAh High Capacity Li-Ion Battery



Key features

- 4800 mAh rechargeable Li-ion battery
- Compatible with ScopeMeter® 190 Series III, 190 Series II, MDA-500 Series Motor Drive Analyzers and 430-II Series Power Quality Analyzers

Product overview: Fluke BP291 4800 mAh High Capacity Li-Ion Battery

The Fluke BP291 Li-ion Battery is a 4800 mAh high-capacity battery for use in Fluke 190 Series II and Series III ScopeMeter Test Tools, MDA-500 Series III Motor Drive Analyzers and 430-II Series Power Quality Analyzers. BP291 High-Capacity Li-ion batteries include a charge-status indicator so you can quickly and easily verify charge status in the field, and may be charged either in the instrument itself, or using the option <u>EBC290 External Charger</u>.

Specifications: Fluke BP291 4800 mAh High Capacity Li-Ion Battery

Specifications*	
Battery	Li-ion, 10.8 V, 5000 mAh, 54 Wh configuration 3ICR19/66-2
Safety	Pollution degree 2
Operating temperature	0 °C to +50 °C (charging may slow or stop at environmental temperatures >40 °C)



Storage temperature	-20 °C to +60 °C
Recommended storage temperature	<35 °C
Humidity	0 °C to 10 °C noncondensing 10 °C to 30 °C @ 95 % (±5 %) rh 30 °C to 40 °C @ 75 % (±5 %) rh
Altitude	3000 m
Weight	330 g
Dimensions	22 mm x 150 mm x 59 mm



Ordering information



Fluke BP291

4800 mAh High capacity Li-ion battery for Fluke 190 seriesII



$\textbf{Fluke}. \ \textit{Keeping your world up and running}. \\ \textbf{\textcircled{\$}}$

Fluke Corporation

PO Box 9090, Everett, WA 98206 U.S.A.

For more information call: In the U.S.A. (800) 443-5853 In Canada (800) 36-FLUKE From other countries +1 (425) 446-5500

www.fluke.com

©2023 Fluke Corporation. Specifications subject to change without notice. 04/2023

Modification of this document is not permitted without written permission from Fluke Corporation.